

Cira Tilt Instruction Manual

Product Code: 7-CIRA

Please carefully read instructions before using the bed.

Alrick.



INTRODUCTION

This instruction manual contains information on the instructions for use and explanations of the signs and symbols used. These instructions for use contain instructions for operating the medical care bed.

These instructions for use may contain inaccuracies or printing errors. The information provided here will be updated periodically and changes due to product maintenance will be included in later editions. Changes or improvements are possible at any time without prior notice. If you have any questions, please contact our customer service team.

The instructions for use must be read and used by every person who operates the medical bed.

PURPOSE

The intended purpose of the Alrick Cira Tilt Bed is to be used for patients in hospital and healthcare facilities. These beds are electrically powered to allow the bed to be placed in various positions suitable to the patient. It is a medical device intended for human medical use only.

The Alrick Cira Tilt Bed moves on castors and features infinitely adjustable motorised height settings. The bedding area can be adjusted by multiple motors, allowing for positions such as sitting or exiting position.

The intended user of the Alrick Cira Tilt Bed is any individual who has a disability or a health condition OR any bariatric (but not limited to) individual who has a disability or a health condition.

The bed should only be used by individuals trained or experienced in proper handling. Users must be instructed on its correct use and familiarize themselves with these instructions. Common users include nursing staff, physicians, therapists, family members, technicians, residents, and cleaning staff.

SYMBOLS EXPLAINED



Manufacturer



For Indoor Use (house)



Type B Application Part
(classification according to the degree of protection against electric shock)



Product with a Thermofuse



Safety Isolating Transformer




Device Protection Class II
(classification according to the type protection against electric shock)



Demko Approval

PRODUCT SERIAL NUMBER PLATE



Serial Number Plate

SERIAL NUMBER PLATE EXPLAINED

- 1 ARTG Number
- 2 Company Logo
- 3 Adult Identification (minimum weight, minimum height, minimum BMI)
- 4 Maximum Occupant Weight
- 5 Maximum Safe Working Load
- 6 Read Instruction Manual - Follow Instructions For Use
- 7 Product Code
- 8 Product Serial Number
- 9 Manufacture Date
- 10 CE Marking
- 11 Type B Application Part
(classification according to the degree of protection against electric shock)
- 12 Switching Power Supply Unit
- 13 Device Protection Class II
(classification according to the type protection against electric shock)
- 14 For Indoor Use Only
- 15 Manufacturer Address
- 16 Company Service Email Address
- 17 Company Phone Number

PRE OPERATION CHECK

WITH BED IN POSITION, CHECK THE FOLLOWING:

Electric beds should ONLY be used where earth leakage is fitted to the power source.

- Check power cord is plugged into 240V supply and is active.
- Check power cord is clear of castors and any moving part of bed.
- Ensure power cord has enough slack so cord is not strained when the bed is raised and lowered.
- Position bed at least 50mm from the floor in the lowest position. Lock castors. This position is critical so as to not damage the wall when raising the bed.
- Before using bed, perform all functions to ensure they are working properly. If not please call your service technician or call Active Medical Technician Team on 1800 282 001

** When fitting safety sides to this bed, please beware of the possible crushing points. The clamps can create crushing points on the rails. **

CIRA TILT MODEL'S AVAILABLE

Cira Tilt Single (900 mm)

Cira Tilt Double (1370 mm)


Cira Tilt King Single (1070 mm)

Cira Tilt Queen Double (1500 mm)

Cira Tilt Extra Wide (1200 mm)

Cira Tilt King Double (1800 mm)

SPECIFICATIONS

	Single	King Single	Extra Wide	Double	Queen Double	King Double
Height Range (mm)	360 - 800	360 - 800	360 - 800	360 - 800	360 - 800	360 - 800
Backrest Angle (degrees)	77	77	77	77	77	77
Kneebreak Angle (degrees)	40	40	40	40	40	40
Trendelenburg Angle (degrees)	16	16	16	16	16	16
Reverse Angle (degrees)	16	16	16	16	16	16
SWL (kg)	300	300	300	300	300	300
Overall Dimensions (mm)	2060 x 1050	2060 x 1220	2060 x 1350	2060 x 1520	2060 x 1650	2060 x 1950
Sleeping Surface (mm)	2000 x 900	2000 x 1070	2000 x 1200	2000 x 1370	2000 x 1500	2000 x 1800
Mattress Size (mm)	1960 x 870	1960 x 1050	1960 x 1200	1960 x 1350	1960 x 1480	1960 x 1780
TGA Number	ARTG 192062					
Manufacturer 	Active Medical, 104 Loughnan Street, Lake Cargelligo, NSW 2672					

ELECTRICAL DATA

U In	230V ~, 50Hz
I In	Max. 1.5A
U Out	24V ... , max. 70W
Op.	10%, Max. 2min./18min.
Protection Class	1Px4
Insulation Class	II Type B
Noise	48DBA



For indoor use (house)



Safety isolating transformer



Type B equipment, as per EN60601+1



Device Protection Class II



Product with a thermo fuse



Demko approval

RE-Tuv approval

IMPORTANT

- Check electric cords regularly and notify an electrician or Active Medical on 1800 282 001 immediately. If any damage occurs cease to use and remove mains power until checked and repaired.
- No wiring is to be modified without authorisation from the manufacturer.
- Ensure the small extension cable is correctly connected to the bed socket and to the 240 volt supply when the bed is in a stationary position. A small light on the control box will illuminate showing a power supply.
- Low Low model beds should be left in the lowest position when a nurse is not in attendance.
- Nurses, cleaners or anyone using Alrick Electric Beds must make sure that the main power cord is clear of any moving parts on the bed before turning the power on.
- Electric Beds should ONLY be used where earth leakage is fitted to the power source.
- When patient condition could lead to patient entrapment the mattress support platform should be left in the flat position when unattended.



OPERATOR CONTROLS



OPERATOR CONTROLS:


Electric Beds are controlled by a handset, which is plugged into the connection port. (Normally located below and at the edge of the mattress base towards the head end).

Handset Code: HB8X255-00



UPPER POSTURE, ELECTRIC BACKREST:

The top pair of push buttons control the electric upper posture as indicated on the handset. Beware of potential pinch points when lowering backrest.



HI-LOW OPERATION:

The second pair of buttons on the handset control the smooth operation of raising & lowering the bed. When raising and lowering the bed, ensure that no person or part of his or her body is under the bed. Failure to do so could cause serious physical injury.



ELECTRIC KNEEBREAK:

The third pair of push buttons on the handset control the operation of the kneebreak. Beware of potential pinch points when lowering kneebreak.



CHAIR POSITION:


The fourth set of buttons operate the chair function.



ELECTRIC TRENDELENBURG:

The last pair of push buttons on the handset control the operation of the tilt.

Beware of potential pinch points when using the tilt function.



NIGHT MODE:

Night Mode returns all functions to the lowest position.

WARNING: It is the responsibility of the carer to make sure there is nothing under the bed. Once checked, they can lower the bed to its lowest point of 360mm.



UNDER BED LIGHT:

The last button on the bottom left controls the under bed light

OPTIONS & EXTRAS



BATTERY BACK UP:

This allows you to safely change the patients position during a power shortage.

CASTORS



STANDARD CASTORS:

Brakes should only be released when the bed is moved.

Red pedal castors: total brake

Green pedal castors: directional.



CENTRAL LOCKING CASTORS:

Brakes should only be released when the bed is moved.

To brake: press lever down.

To direction: lift lever all the way up.

To free wheel: centralise lever.

OPERATING THE RAILS WARRANTY: 12 months

Special care and caution are required when using siderails. Siderails are used primarily as patient protection. Under no circumstances may the siderails be used as a means of patient immobilization. When using siderails, the suitability must be checked taking into account the physical and mental characteristics of the respective resident. In particular, the distances between the bars and stays must be observed in relation to the physique of the resident so that becoming trapped or slipping through can be ruled out. If necessary, additional side protection pads should be used to reduce the gap dimension or close the gaps. This decision is the sole responsibility of the appropriate medical supervisor.

- Incorrectly fixed siderails can sink again, causing the resident to fall. In each new position, shake to make sure that the siderail is properly engaged. Correct engagement can also be checked visually for divided siderails: The yellow detent buttons must be visible in the detent recess of the telescopic mechanism on both sides.
- If necessary, also use an attachment for siderails as an elevation when using raised mattress systems and the available siderail pads (accessories) as an important additional safety and protection feature.
- When lowering or raising siderails, fingers, hands or other parts of the resident's body can be crushed. Raise or lower the siderails with utmost attention. And only grip at the positions intended for this purpose according to the instructions for use.
- Restless and disoriented persons can get caught in the free spaces of the siderails. Protect them by using siderail pad
- The siderails can be overloaded and damaged by large loads, bracing or pulling. Avoid such loads.

The Four-Point siderails are divided into two parts on each side of the bed. There is a release button on the bottom of each of the outer ends of the handrail, which can be used to release the lock. It can be extended in 4 steps up to the height of 400 mm.

Guide the siderail with both hands to ensure even lowering/lifting in both guides.




1. Stage: Accessibility

The siderail is in the lowered position.




2. Stage: Mobilisation

3/4 protection at a height of 247 mm.



3. Stage: Security

Full protection at a height of 349 mm.



3. Stage: Increased Safety

Increased protection at a height of 407 mm with higher mattress thickness

If you want to raise the lowered Four-Point siderail:

- Grasp the Four-Point siderail by the handrail, and lift it until the Four-Point siderail clicks into the next higher position.
- To raise the Four-Point siderail further, simultaneously press in both release buttons and then pull the siderail upwards by the handrail until the Four-Point siderail locks into the next higher position with a clear click.

To lower the Four-Point siderail, reverse the procedure:

- Simultaneously press in both release buttons and lower the Four-Point siderails on the handrail to the desired position.

SAFETY & WARNINGS

- Electric Beds should ONLY be used where earth leakage is fitted to the room or building.
- If any cords are cut or damaged the bed MUST not be used until checked, fixed and tagged.
- No patient, service personnel or any person must lie underneath an Alrick bed while power is turned on, if not obeyed this may result in serious bodily injury or death if the bed controls are actuated.
- Please check no person or object is under the bed before using.
- Hand control to only be given to fully competent patients.
- Never leave hand control unattended with bed switched on unless patient is fully competent.
- Fit optional nurse lockout for safety of incompetent patients.
- Only one person is to be on the bed at any time.

MAINTENANCE

- Beds should be inspected monthly for abnormal wear. For example, bolts loose or missing, castors, rubbers or brakes damaged.
- Rails must be inspected regularly for any wear or movement that may cause injury, do not use if unsure.
- If one or more of the bed functions is found not working correctly, notify the manufacturer for advice.

EVERY 6 MONTHS:

1. Check there is no visible damage to electrical system or mechanical parts.
2. Check wiring is secure and not interfering with bed movement.
3. If actuator shafts are not clean - wipe chromed shaft of actuators to remove lint, grease & dirt.
4. Extended actuators fully. (Do not wash as this removes fine layer of oil).
5. Correct operation of all bed functions including castor locking.
6. Check all accessories including rails are still functioning properly.
7. Check all pivot points on the bed for wear. Replace pivot pins or bolts and removable bushes (where fitted) if a pivot has a significant play.
8. Check all functions still run smoothly.

ANNUALLY:

- Professional service of product to be carried out annually & documented in 'Record of Performed Service Procedures and Regular Service Maintenance in the back of this instruction manual.

CLEANING

- Cleaning can be done often - using domestic detergents.
- Do not use acids, alkaline or any other chemical solvents
- When cleaning actuators, disconnect from mains power before wiping down with a moist cloth.
- When cleaning, be careful of entrapment, as it could cause physical injury.

Control Box



CO715060-00-AU



BA1616121100



2300 Split partner bed

PI2+19200S000000


Actuators

Channel 1
Back
Back Right




LA27-C652-00
or J13703

Channel 2
HiLo Head



400760-00

Channel 3
Leg
Thigh
Leg Right



LA27-C586-00

Channel 4
HiLo Leg




400760-00

Channel 5
Calf
Leg Left



LA27-C586-00

Channel 6
Back Left




LA27-C652-00
or J13703

Handsets



MJB5050117-00



ACOMV0612001+110500



HB8X016-00



Option:
ACC610001+B0000

Option:
ACC610002+B0000

Option:
ACK0003S60C1800

Cables & Accessories

Qty	Item Number	Description
1	SML1912363	Power Cable Flying Earth
2	09680000-1700	Minifit 6 Pin 1700mm CO to Actuator
2	09680000-1100	Minifit 6 Pin 1100mm CO to Actuator
1	0964461-0800	Modjack 800mm CO to MJB
1	0965069-0650	Modjack Ext 650mm Coil MJB to HB
1	1015W1001	Bracket CO/PI2 to Frame
1	1019W7001-00185	Minifit 4 Pin 185mm CO to Battery
1	1015W1001	Bracket CO/PI2 to Frame (PI2 Option)
1	09680000-1700	Minifit 6 Pin 1700mm (PI2 Option)
1	09680000-1100	Minifit 6 Pin 1100mm (PI2 Option)
1	1019W7002-00180	Minifit 4 Pin 180mm CO to PI2 (PI2 Option)



Joint System Drawing

LINAK Spec: BAU01249	System Name: 2300/4000 Multi Purpose System
LINAK Australia	Active Medical

LOCKING AND UNLOCKING LOCK OUT FUNCTIONS (optional)


HB400 HANDSET:

1. Locate the fob over the top of the handset as shown.
2. Then press the down key of the function to be locked out or unlocked. When locked, the LED function will be illuminated.




ACOM ATTENDANT CONTROL:

1. Press and hold the key button.



2. Then press the down key of the function to be locked or unlocked.

RESETTING A CIRA TILT BED FITTED WITH A CO71 LINAK CONTROL BOX:




Press both hi-low buttons (up and down) together

- It will beep for 5 to 10 seconds
- don't let go of the buttons once beeping has stopped.
- After another 10 seconds the bed will start to rise, it will go to the maximum height and return to the minimum height.
- through this process don't let go of the buttons
- Once at the lowest position the back and knee function motor will retract right in, once this is complete, you will get a double beep.
- only let go of the buttons once you have heard these beeps

TROUBLESHOOTING

TROUBLESHOOTING WITH A HB400 HANDSET:



Lock LED flashing on backrest function - issue with either the backrest actuator or cable.

Lock LED flashing on hi-low function - issue with the head end hi-low actuator or cable.

Lock LED flashing on the kneebreak function - issue with the kneebreak actuator or cable.

Lock LED flashing on tilt function - issue with either the foot end hi-low actuator or cable.

If all LED lights are flashing - issue with either handset or the cables connecting it to the control box.

TROUBLESHOOTING WITH ACOM ATTENDANT CONTROL:

If all LED lights are flashing - issue with either handset or the cables connecting it to the control box.

Lock LED flashing on backrest function - issue with either the backrest actuator or cable.



Lock LED flashing on tilt function - issue with either the foot end hi-low actuator or cable.

Lock LED flashing on the kneebreak function - issue with the kneebreak actuator or cable.

Lock LED flashing on hi-low function - issue with the head end hi-low actuator or cable.


PAIRING BLUETOOTH DEVICES (optional)

DIRECT PAIRING

Direct Pairing is used for pairing a LINAK control directly to a LINAK control box that supports BLE.

1. Enter pairing mode
2. Move the hand control closer to the control box you want to pair it with
3. Pair the hand control with the control box

ENTER PAIRING MODE



Place the magnet key over the **Hall Element (O)** and press **Hi-Lo Down Button** for 3 seconds to enter the pairing mode.




When in pairing mode, the control box buzzer will begin to beep, and the LED will begin to blink. The control unit LED will also begin to blink with the same frequency as the nearest control box.



In pairing mode, the light/sound frequency will increase when the HB200 gets closer to a control box.

CONNECTING TO THE CONTROL BOX



Press KEY 1 to connect the hand control box

When the hand control LED is blinking fast and the control box gives a high frequency sound in the same speed, the devices are ready for pairing. To finalise the pairing, press **Hi-Lo Down Button** on the hand control. Then, adjust your application with the handset.

TROUBLESHOOTING

ENGINEER AND SERVICE - FAULT FINDING

PROBLEM	CAUSE	REMEDY
No functions	No power to system	Check the mains and/or battery connections
No functions LED indication ON	Possible loose connection	Check handset/ footswitch connections
Actuator will drive OUT but not IN	Position lost on actuator	Try initialisation procedure, if problem persists, replace actuator
Actuator will drive IN/OUT. Intermittent beeping from control box when button is pressed.	- Cable to actuator disconnected or faulty - Actuator is faulty	- Replace faulty component - Perform reset and initialisation procedure
One or more actuators stop while driving	Actuator is overloaded	Reduce or remove load on chair and try again
HILO columns will not drive IN/OUT	- Cable to column disconnected or faulty - Column is faulty	- Replace faulty component - Perform manual reset procedure

OPEN BUS DO'S AND DON'TS

1. Before connecting or disconnecting OPENBUS™ accessories or handsets to/from the system:
 - Disconnect the CB from mains power.
 - Wait at least two minutes after the last key press.
2. Only connect the specified type of Actuator to the appropriate port on the CB, refer to section 'System Overview' and section 'Assembly'.
3. Only connect the specified handset to the system, refer to section 'Assembly'.
4. Wait at least 2 minutes after the last key press before removing the BA18xx battery from the system. Otherwise battery operation may be blocked for 30 minutes.
5. Do not connect/disconnect an actuator while pressing any keys on the handset or foot switch.
6. Use O-rings on all connectors and secure all connectors/cables with cable locking devices supplied.
7. Use blind plugs with O-rings to seal any unused ports on the CB and junction boxes.
8. Do not stretch or unduly strain any of the cables and connectors.

RECORD OF PERFORMED SERVICE PROCEDURES AND REGULAR SERVICE MAINTENANCE

RECORD OF PERFORMED SERVICE PROCEDURES AND REGULAR SERVICE MAINTENANCE

RECORD OF PERFORMED SERVICE PROCEDURES AND REGULAR SERVICE MAINTENANCE

Date

Purpose and Description

Performance

RECORD OF PERFORMED SERVICE PROCEDURES AND REGULAR SERVICE MAINTENANCE

Date	Purpose and Description	Performance

I confirm that technical staff have undertaken the repairs as described above and the equipment is now functioning properly.

Date _____ Customer Signature _____

ALRICK CIRA TILT BED WARRANTY

CONDITIONS

Your Alrick Cira Tilt Bed is guaranteed for 10 years on the frame from the date of purchase, 5 years on the electrics (LINAK), 1 years on castors, and 1 year on the battery. Professional service of product to be carried out annually and documented in the 'Record of Performed Service Procedures and Regular Service Maintenance. This warranty applies only on the parts manufactured or approved by Active Medical.

EXCLUSIONS

This warranty does not extent those items that may need replacement due to normal wear and tear, or to damaged caused by misuse or accident, for which Active Medical or it's distributors cannot be held responsible.

If your bed should need attention due to failure as defined under the warranty above, please contact the distributor from whom you purchased your bed.

Please fill in the below and retain for your own records.

Date Purchased

Model

Serial Number(s)

Distributor

Distributor Address

Distributor Contact Number

ACCESSORIES



Bed Rail Protector (3983)



Bluetooth Handset (7-BT-4-HB200)

Warranty: 12 months



IV Pole

Warranty: 12 months



T-Bar Buffer

Warranty: 12 months



Self Help Pole

Warranty: 12 months



Reading Light

Warranty: 12 months



UPS Emergency Bed Power Unit

Warranty: 12 months



Mattresses

RV 140825

Alrick.

1800 282 001
hello@alrick.au
www.alrick.au